

ESD Survey Kit Operation and Maintenance



Made in America



Description

The Charleswater Item 99200 Survey Kit contains everything needed to conduct an ESD Control Program survey, per EN 61340-5.

The ESD Survey Kit is your "appropriate measuring equipment" specified by EN 61340 5-1 paragraph 10.1 requiring "periodic audits of facilities".

Satisfy EN 61340 5-1 paragraph 9.8 six-monthly checks requirement. Use the Digital Field Meter to measure electrostatic fields and use the tester to electrically test ESD garments resistance and panel to panel conductivity.

The ESD Survey Kit can be used to fulfill testing required by EN 61340-5-1 paragraph 10, including resistance and charge generation of existing working surfaces, flooring, seating, clothing and racking or shelving, protective packaging materials and transit boxes.

This clearly states the necessity of having the means to perform regular surveys at your facility. The 99200 will enable you to do this.

Contents of the Survey Kit

Corresponding technical document numbers are noted as well as web site links.

- (1) 99091 Ionisation Test Kit (PPE-5048.E)
<http://www.charleswater.co.uk/pdf/PPE-5048.E.pdf>
- (1) HBM Test Fixture
<http://www.charleswater.co.uk/pdf/PPE-5001.E.pdf>
- (1) 99105 Surface Resistance Test Kit (PPE-5075.E)
<http://www.charleswater.co.uk/pdf/PPE-5075.E.pdf>
- (1) Fluke 10 Meter
- (1 set) Training Paddles
- (1) 60148 Wrist Band (PPE-5001.E)
- (1) 60341 Coil Cord (PPE-5001.E)
<http://www.charleswater.co.uk/pdf/PPE-5001.E.pdf>
- (2) 70004 Foot Grounders (PPE-5006.E)
<http://www.charleswater.co.uk/pdf/PPE-5006.E.pdf>
- (1) Heavy duty carrying case

Limited Warranty

Charleswater expressly warrants that for a period of one (1) year from the date of purchase, Charleswater ESD Survey Kits will be free of defects in material (parts) and workmanship (labour). Within the warranty period, the product will be tested, repaired, or replaced at Charleswater's option, free of charge. Call our Customer Service Department at 00 44 (0) 1892-665313 for a Return Material Authorisation (RMA) and proper shipping instructions and address. Include a copy of your original packing slip, invoice, or other proof of purchase date. Any unit under warranty should be shipped prepaid to the Charleswater factory. Warranty repairs will take approximately two weeks.

If your unit is out of warranty call Customer Service at 00 44 (0) 1892-665313 for a Return Material Authorisation (RMA) and proper shipping instructions and address. Charleswater will quote repair charges necessary to bring your unit up to factory standards.

Warranty Exclusions

THE FOREGOING EXPRESS WARRANTY IS MADE IN LIEU OF ALL OTHER PRODUCT WARRANTIES, EXPRESSED AND IMPLIED, INCLUDING MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE WHICH ARE SPECIFICALLY DISCLAIMED. The express warranty will not apply to defects or damage due to accidents, neglect, misuse, alterations, operator error, or failure to properly maintain, clean or repair products.

Limit of Liability

In no event will Charleswater or any seller be responsible or liable for any injury, loss or damage, direct or consequential, arising out of the use of or the inability to use the product. Before using, users shall determine the suitability of the product for their intended use, and users assume all risk and liability whatsoever in connection therewith.